Certificate of Calibration



This calibration certificate certifies that the equipment enclosed conforms to DIN 875 Grade 00

Accuracy for squares according to DIN 875	
Len <mark>gth (</mark> mm)	Deviation of right angle according to accuracy classes (mm)
	00
150	0.004
250	0.005
350	0.006

The measurements have been made using equipment traceable to national standards.

Signed ..

Andrew Wiltshire (Quality Systems Engineer)

Villshire.

Manufactured in Axminster using the latest CNC production techniques.

Axminster Tools, Weycroft Avenue, Axminster, Devon EX13 5PH.

Call 03332 406406. Visit axminstertools.com

